



Spinal

Conform Maternity Belt			
CB/900	CB/960	CB/1020	CB/1080
CB/1160	CB/1260	CB/1360	CB/1460

Description

An orthosis designed to relieve backache and fatigue. The brace is 200mm in depth and features a wide elastic that passes underneath the stomach, alleviating the weight on the lower spine and lifting the belly. The touch and close hook and loop fastenings adjust as the bump grows. The belt contains latex but woven into the material in a way so that the latex does not come into contact with the skin directly.

Intended Purpose

To provide back and pelvic support for an array of injuries and conditions, including Pre or Postoperative treatment for: Pubic support, Backache and fatigue, Sacroiliac support throughout all stages of pregnancy.

Sizina

Measure the circumference of the hips.

⚠ Fitting Instructions

We recommend the initial fitting of the device be conducted by a suitably qualified Healthcare Professional who will advise fitting, removal and period of use.

- 1. Open the front hook and loop fastening See Fig.1.
- 2. Align the belt to the midline of the lower back, with the padding lying in the middle over the spine See Fig.2.
- Wrap the elastic around the front underneath the stomach, to alleviate the weight on the lower spine and lift the belly See Fig.3.
- 4. Fasten the hook and loop fastening See Fig.4.









- Must not be used by individuals for whom compression is contraindicated
- · Do not use on open wounds
- · Any known allergy and/or hypersensitivity to any material listed in the product composition
- · Caution should be used if prescribed to patients with diabetes, vascular deficiency and neuropathy

Marnings and Precautions

- · Carefully read all instructions and warnings before use
- Follow all instructions to ensure proper performance of the brace
- Do not use if liniments, ointments, gels, creams or any other substances have been applied to the affected area
- · Do not reuse for another patient, doing so risks cross-infection and can compromise product integrity
- To ensure the straps are not over tightened, perform regular skin and circulation checks, especially for patients with diabetes, vascular deficiencies and neurological conditions
- Should any adverse reactions occur, please remove, do not use and contact your healthcare professional or provider
- · Check brace and all components for wear and tear
- · The hook and loop fasteners must always be fastened when the device is not being worn or when being washed
- The durability of the device may be compromised by certain factors, e.g. objects with sharp edges or damage to the hook and loop fasteners

Mashing and care instructions

- · Cover all the hook and loop straps
- Hand wash the brace in cool water using a mild detergent
- · Rinse and air dry flat
- Do not spin or tumble dry
- Do not use fabric conditioner at any time
- · Check for any degradation that may affect use
- Periodically check the hook and loop straps are operating securely

Product composition

Elastic, Hook and Loop, Latex

Storage and transport conditions

Store in a cool, dry place out of direct sunlight, in the original packaging.

Recycling and disposal

Packaging and constituent parts should be recycled or safely disposed of in accordance with local or national laws.

Serious incident

Report any serious incident to the manufacturer and the competent authority of the EU Member State, or Country in which you reside.

Wash symbols

Hand wash

Dry flat

Do not tumble dry

Do not bleach

Do not dry clean

Do not iron

Do not tumble dry

Contains natural rubber latex



Beagle Orthopaedic

Prospect Building, Duttons Way, Shadsworth Business Park, Blackburn, BB1 2QR e: info@beagleortho.com t: +44 (0) 1254 268 788 w: beagleorthopaedic.com



This Medical Device conforms to the requirements of Regulation (EU) 2017/745 of the European Parliament and of the council of 5 April 2017

EC REP

International Associates Auditing & Certification Limited

The Black Church, St. Mary's Place, Dublin, D07 P4AX